





Siargo Ltd.



MF5200 Series with Detachable Display

MEMS Oxygen Flow Meters

SIARGO MEMS FLOW SENSING PRODUCTS

User Manual (VB.4)



© 2018 Siargo Ltd.



Siargo Ltd.

MEMS Oxygen Flow Meters

MF5200 Series

User Manual

Document No. 04-2018-M4 EN

Issue date: 2018.04

Revision: VB.4

Siargo Ltd.

3100 De La Cruz Boulevard,
Suite 210,
Santa Clara, CA 95054
USA

Tel: +1(408)969.0368

Email: info@siargo.com

© Copyright 2018 by Siargo Ltd.

Siargo Ltd. and its subsidiaries reserve the rights to change the specifications and/or descriptions without prior notice. For further information and updates, please visit:

www.Siargo.com

RESTRICTION ON USE

1. This meter is manufactured for hospital oxygen metering. Do not alter any hardware and software of the product. Any modifications might cause damage and unexpected events.
2. All practices for electronic device safety should apply.
3. Do not use this product in any environments where human safety may be at risk.
4. Only a qualified person from Siargo or a person who is accredited by Siargo can perform troubleshooting services to the product, Siargo is otherwise not liable for any consequences thereafter.

SAFETY PRECAUTION

1. The product can be utilized to measure and/or monitor in-line mass flow rate of any clean, dry and preferably gases with constant concentration in industrial applications. For other special gases or variable concentration gases, the product may not function properly or even can be damaged. Please contact Siargo for further information.
2. The operational environments of the product are illustrated in the section of product specifications. If the product is used for other circumstances, the product may not function properly or even can be damaged.
3. Operation, installation, storage, and maintenance of the product must strictly follow the instructions illustrated in this user manual. Otherwise, unpredicted damage and even injuries or other severe situations could be induced. All the installation, storage, and maintenance of the product must be performed by skilled workers. This user manual should be placed near the product for easy access.
4. Before using the product, the user should read this user manual completely and in details so that the user is well understand all the important instructions.
It is recommended that the product should be re-calibrated and serviced in every two years or at a time of desire.

Contents

RESTRICTION ON USE	1
SAFETY PRECAUTION	1
Contents	2
1 Overview	3
2 Models and Selection	4
3 Product Description	4
4 Specifications	5
5 Installation	6
5.1 Physical Dimensions	6
5.2 Installation Instructions	6
5.3 Attentions	6
6 Operation and Communication	8
6.1 Cable Definition	8
6.2 LCD Display	8
6.3 RS485 Communication	8
6.4 Setup via Buttons	9
7 Safety and Maintenance	11
7.1 Wetted Materials and Compatibility	11
7.2 Safety Precautions	11
7.3 Maintenance	11
8 Warranty	12
9 Customer Service and Order information	13

1. Overview

MF5200 series oxygen flow meters are specially designed for oxygen metering system in a hospital.


These meters are operated with the standard industrial user interfaces (RS485), and user defined alarm functions for better monitoring or control over the network. The off-the-shelf products (MF5212 and MF5219) are for gas flow measurement of 0 ~ 300 SLPM and 0 ~ 800 SLPM, respectively.

MF5200 series oxygen flow meters feature:


- ♦ Integrated MEMS mass flow sensors and pressure sensors
- ♦ Detachable display that can be placed away from the flow channel via a specially designed cable
- ♦ LED display for both instant flow rate and flow accumulation
- ♦ Standard RS485 for remote data communication
- ♦ Hygienic stainless steel body ready for oxygen metering

Working Principle and Package


The MEMS calorimetric sensor is installed at the flow channel wall forming a plate that serves as an additional flow conditioner from the boundary layer configuration resulting in a laminar flow. The mass flow measurement is established as the fluid carries heat away from the heater causing the redistribution of the temperature field. Accurate flow rate is obtained by calibration with the standard fluid at the preset conditions.



Time-averaged velocity profile boundary layer



2. Models and Selection



* There is flow rate number only.


Typical flow range:

Model	L/min	m ³ /h
MF5212	200, 300	12, 18
MF5219	600, 800	36, 48


3. Product description

The parts are illustrated as below:

Meter Body



Display



4. Specifications

Model		MF5212	MF5219	Unit
Flow	Max. Flowrate	200, 300	600, 800	SLPM
	Min. Flowrate	0.3	0.3	SLPM
	Turn-down ratio	50:1		
	Accuracy	±(2.0+0.5FS)		%
	Repeatability	±1.0		%
Pressure	Output Range	15 ~ 700		kPa
	Repeatability	±2.5		%
Power Supply		8~24 Vdc (With AC adapter)		
Output		RS485 (Modbus)		
Display		LED (Detachable)		
Display Unit		Instant flowrate: m ³ /h or L/min, Flow accumulation: m ³		
Display Resolution		Instant flowrate: 0.1L/min, Flow accumulation: 0.001m ³		
Keyboard		3 Keys		
Max. Pressure		1.0		MPa
Storage Temperature		-20~+60		°C
Operating Temperature		-10~+55		°C
Humidity		<95%RH(No icing or condensation)		
Electrical Connection		Inputs/outputs Cable; Detachable LED Cable		
DN		12	19	mm
Mechanical Connection		DN 12 tri-clamp with M22*1.5 Connector	DN 19 tri-clamp with M33*1.5 Connector	
Weight	Meter Body	1450	1600	g
	Display A	170		
	Display B	260		
Calibration		Air @ 20°C, 101.325kPa		
Protection		IP40		

5. Installation

5.1 Physical Dimensions




Figure 5-1

5.2 Installation Instructions

The product at the time of shipment is fully inspected for product quality and meets all safety requirements. Additional safety measures during the installation should be applied. This includes, but is not limited to leakage verification procedures, standard EDS (electrostatic discharge) precautions, DC voltage precautions, and heavy duty precautions. Other tasks such as calibration, part replacement, repair, and maintenance must only be performed by trained personnel. Upon requests, manufacturer will provide necessary technical support and/or training of the personnel.

Do not open the product cover or alter any part of the product. Any such actions will forfeit the terms of the warranty and cause the liability to any damages thereafter.

The product is preferably to be installed horizontally. Flow direction should be aligned with the arrow mark on the meter body. If the flow fluid may have particles or debris, a filter is strongly recommended to be installed upstream of the meter.

Please follow the following steps to complete the installation:

- a) Upon opening the package, the product physical integrity should be inspected to ensure no visual damage.
- b) Before installation of the product, please ensure that the pipe debris or particles or any other foreign materials are completely removed.
- c) Cautions during installation:
 - (I) It is preferably to first install the inlet end of the meter and then the outlet end of the meter; To ensure the measurement accuracy, an upstream straight pipe of length no less than 10DN and a downstream straight pipe of length no less than 5DN should be in place. (Fig 5-2)
 - (ii) During installation, please make sure no any foreign materials (such as water, oil, dirty, particles, etc.) falling into the pipe.




Figure 5-2

(iii) Straight pipe requirement of special cases.




Figure 5-3

- d) Connect electrical wires for LCD, and then electrical wires for inputs/outputs. Please pay special attention to power supply range (i.e., +8~+24 VDC) and power supply polarization (see the description on Electrical Interfaces in this manual).
- e) When connect the communication wires, please make sure that the wires are correctly connected to the proper ports on your data device/equipment.
- f) Turn on the power supply, and make sure that the LCD works correctly.
- g) Slowly open the valves at the both ends of the pipeline, and the meter should then start to measure the flow in the pipeline
- h) Completion of the installation.

5.3 Cautions

- a) Don't try to loose any build-in part of the product.
- b) Ensure electrical wires for the inputs/outputs to be reliably connected.
- c) Release all the installation stresses so that no stresses will be exerted on the product.
- d) The product should avoid strong electromagnetic interference sources nearby or periodic mechanical shocks to its body or pipeline.
- e) Slowly open/close valves to prevent abrupt pulse flow impact, which may damage the product.

6. Operation and Communication

6.1 Cable Definition

The electrical interfaces are defined as below:

Color	Definition
Red	Power Supply (+8 ~ +24 VDC)
Black	Power GND
Green	RS485 (A)
Brown	RS485 (B)
Violet	Alarm High Signal Output
Yellow	Alarm Low Signal Output
White	Alarm Signal GND



Figure 6-1. Accessory Cable 1 for Inputs/Outputs (Part number: IC7B-100, Length: 100 cm)



Figure 6-2. Accessory Cable 2 for Detachable LED (Part number: IC8B-100-IC8B, Length: 100 cm)

6.2 LED Display

Normally, the LED display looks as Fig 6-3:




Figure 6-3a. Normal Display (Total & Flow rate)




Figure 6-3b. Normal Display (Pressure & Flow rate)

6.3 RS485 Communication

For purposes of computer control and networking, the RS485 is used for communication with the following settings:


Baud rate (Bits per second):	57600(Single-device communication) 9600(Multi-device communications)
Date bits:	8
Stop bits:	1
Parity:	None
Flow control:	None

6.4 Setup via Buttons

6.4.1 Button definition

Three buttons:   

: Selection/confirmation of a setting

: Scroll up the setup menu

: Scroll down the setup menu

6.4.2 Operation

(1) The user interface (Figure 6-4):





Figure 6-4 User Interface

Button  is used for function selection. After press it, the menu asks for password (authentication mode).

(2) Password interface (Figure 6-6):

In the password menu, the flow measurement will not be interrupted, whereas the first line of the LED display will show the password menu as Figure 6-5:







Figure 6-5 Password menu

- The password consists of six numeric digits. The blinking digit can be assigned a numeric value, which can be selected from 0-9 through the up/down buttons (/).
- After selecting a desirable value, press  to conform the selection, and the proceed to the next digit.
- After the password is correctly set, the meter enters the function setup menu. Otherwise, the meter returns back to the user mode. (**NOTE: The default password is 111111**)



(3) Function setup menu:

▲ Caution: If you want change any settings, please refer to the manual, otherwise the meter maybe work abnormally.

A.

“qUIT”, exit from the setup mode (this is the default option).

B.

“CLr ACC”, reset the flow accumulation reading to zero. Press  and “nO” or “yES” will be display as below, press  to confirm and exit.

C.

“OFFSEr”, reset the offset of the meter. Press  and “nO” or “yES” will be display, press  to confirm and exit.

Notes: Please ensure there is no flow in the pipeline and then reset the offset.

D. **SET PASS**

“SET PASS”, Set the password.

Notes: Please remember the new password and placed it properly.

E. **SET Addr**

“SET Addr”, Set the address for *Multi-device communication mode*. The address will be shown as below:

--006--

F. **PrOTOCOL**

“PrOTOCOL”, select *Single-device communication* or *Multi-device communication*.

- Press **(M)**, the display will show the protocol menu; if the value is “000”, the meter is working in *Single-device communication mode*; if the value is a number between 001 and 255, the meter is working in *Multi-device communication mode*. (e.g, if the value is 006, it means the meter is working in Multi-device communication mode and the address is 006.)
- Press **(◀/▶)** to switch in two communication modes.
- After selection, press **(M)** to confirm and exit.

--006

G. **UnITTYPE**

“UnITTYPE”, Set the unit of instant flowrate.

- Press **(◀/▶)** to switch in two units.
- After selection, press **(M)** to confirm and exit.

--SL-- --n3--

H. **SET GCF**

“SET GCF”, Set the GCF (Gas Correction Factor), default value is 1000.

GCF 1000

I. **ALAr HI**

“ALAr HI”, Set the alarm of maximum instant flowrate, default value is 65535.

It is signal low (0V), when instant flowrate is larger than “ALAr HI” value, it will be signal high (3V).

HI -65535

J. **ALAr LO**

“ALAr LO”, Set the alarm of minimum instant flowrate, default value is 00000.

It is signal low (0V), when instant flowrate is smaller than “ALAr LO” value, it will be signal high (3V).

LO -00000

K. **InTERvAL**

“InTERvAL”, Set the response time of the meter, selectable: 125, 250, 500, 1000, 2000 & 4000 ms.

L. **PrS dISP**

“PrS dISP”, Set the pressure display modes, default value is 00000. See detailed in next page.

---0---




Figure 6-6 Buttons operation

(4) Communication Modes switches

A. From *Single-device communication* to *Multi-device communication*

- Set the address of the meter (value of `SEr Addr`) as (3).E. such as 006;
- Set the meter to *Multi-device communication mode* (select value of `PrOfOCoL`) as (3).F.

B. From *Multi-device communication* to *Single-device communication*

- Set the meter to *Signal-device communication mode* (select value of `PrOfOCoL`) as (3).F.

7. Safety and Maintenance

7.1 Wetted Materials and Compatibility

The meter body and pipe are made of 304 stainless steel. Sensors comprise of silicon, silicon nitride and silicon dioxide and the sensor surfaces are passivated with silicon nitride and silicon dioxide. The electronic sealing is provided by RTV (room temperature vulcanizing) silicone sealant WR-704 composed of HOCH₃(SiO)_nCH₃H.

7.2 Safety Precautions

The product is designed for use with general purpose gases such as air and nitrogen. It is advised that the products are best used for non-explosive clean gases. The sensors cannot be used for gas metrology of fluoride or fluoride containing gases. For updates of the product certification information, please contact manufacturer or visit www.Siargo.com. Use for other gases such as extreme corrosive and toxic may cause the product malfunctioning or even severe damages. The product sealing is ensured to work under working pressure of 1.0 MPa and is leakage proof before the shipment. But cautions and further leakage test are important at installation as well since any leakage could cause severe safety issue. The power supply for this product is 8~24 VDC, all precautions and measures for electrical voltage handling must apply.

Attention: any alternation and/or improper use of the product without the permission of the manufacturer can cause unpredicted damages and even injuries or other severe situations. Siargo Inc or any of its employees, subsidiaries shall not be hold and indemnified against such consequences due to such circumstances via improper use of the product.

7.3 Maintenance

Attention: without prior permission of the manufacturer, please do not attempt to alter any parts of the product as it may cause unrecoverable damages. If there are questions or doubts, please contact manufacturer immediately before further actions. Please ensure the DC power is off before disassembling the sensor.

All maintenance of the sensor should be done by trained and certified personnel by Siargo Ltd.

8. Warranty

Siargo warrants the products sold hereunder, properly used and properly installed under normal circumstances and service as described in the user's manual, shall be free from faulty materials or workmanship for 180 days for OEM products, and 365 days for non-OEM products from the date of shipment. This warranty period is inclusive of any statutory warranty. Any repair or replacement serviced product shall bear the same terms in this warranty.

Siargo makes no other warranty, expressed or implied and assumes no liability for any special or incidental damage or charges, including but not limited to any damages or charges due to installation, dismantling, reinstallation or any other consequential or indirect damages of any kind. To the extent permitted by Law, the exclusive remedy of the user or purchaser, and the limit of Siargo's liability for any and all losses, injuries or damages concerning the products including claims based on contract, negligence, tort, strictly liability or otherwise shall be the return of products to Siargo, and upon verification by Siargo to prove to be defective, at its sole option, to refund, repair or replacement of the products. No Action, regardless of form, may be brought against Siargo more than 365 days after a cause of action has accrued. The products returned under warranty to Siargo shall be at user or purchaser's risk of loss, and will be returned, if at all, at Siargo's risk of loss. Purchasers or users are deemed to have accepted this limitation of warranty and liability, which contains the complete and exclusive limited warranty of Siargo, and it shall not be amended, modified or its terms waived except by Siargo's sole action.

This warranty is subject to the following exclusions:

- a). Products that have been altered, modified or have been subject to unusual physical or electrical circumstances indicated but not limited to those stated in the user's manual or any other actions which cannot be deemed as proper use of the products
- b). Siargo does not provide any warranty on finished goods manufactured by others. Only the original manufacturer's warranty applies.

9. Customer Service and Order information

Siargo Ltd. is making every effort to ensure the quality of the products. In case of questions, and or product supports, please contact customer service at the address listed below. We will respond your request in a timely fashion and will work with you toward your complete satisfaction.

Customer service and all orders should be addressed to

Headquarters

Siargo Ltd.
3100 De La Cruz Boulevard, Suite 210,
Santa Clara, California 95054, USA
Phone: +01(408)969-0368
Email: info@Siargo.com

Representative in Japan

Marubeni Information Systems Co., Ltd.
Device Solutions Department
14th Floor, Shinjuku Garden Tower,
3-8-2, Okubo, Shinjuku-ku, Tokyo 169-0072, Japan
Phone: +81-3-4243-4160
Fax: +81-3-4243-4198

Representative in Europe

IDENTIC GmbH
In der Siedlerruh 24
69123 Heidelberg / Germany
Phone: +49-(0)6221-7509777
Fax: +49-(0)6221-7509779
Email: info@identic.de

For orders, please provide accurate and full post address. Siargo will not ship to P.O. Boxes or via a third party.

For further information and updates, please visit www.Siargo.com.

Appendix: Revision History

Revision B.4 (April 2018):

- ✎ Updated the *Function button (Fun)* to *menu button (M)*.

Revision B.3 (August 2017):

- ✎ Correct the LCD photo (Cover);
- ✎ Corrected the representative in Europe information (9. Customer Service and Order information);
- ✎ Added the revision history (Appendix).